

^{224}Ra standardization: radiopharmaceutical perspective

International Workshop on Standards and
Measurements for Alpha Emitting Nuclides in
Therapeutic Nuclear Medicine

22 - 23 February 2024



Radspherin



- **Current status:** clinical phase 2b trials - mid-2024
- Produced and analysed in **Oslo, Norway**
- Delivered ready-for-use to several sites in **Norway, Sweden, Belgium, Spain**, and expanding.





Gro Hjellum
Chief Operating Officer



Roy Hartvig Larsen, PHD
Chairman



Øyvind Bruland, Prof. emerit.
MD, PhD
Board member




Patient


Clinics

Quality
control


Production



Clinics



Patient



Clinics



- 1) The dose **calibrators on clinical sites can be of different** models and vendors (Atomlab from Biodex, IBC-Lite/VDS from Comier, VIK from Veenstra, etc.). They are diverse, but very typical.
- 2) **Dial settings in clinics are set under Oncoinvent supervision.** According to our standard, “All affirmative measurements should be found within 5% of the decay corrected activity”
- 3) In practice, the results are within 2,5%.

Quality control



Calibration standard



Clinics



Dose calibrator dial setting for Radspherin®

Table 2. Decay Correction Table

Time from reference time (days: hours)	Decay Correction Factor	Time from reference time (days: hours)	Decay Correction Factor
0 days: -4 hours	1.02	1 days: +1 hours	0.820
0 days: -3 hours	1.02	1 days: +2 hours	0.813
0 days: -2 hours	1.02	1 days: +3 hours	0.807
0 days: -1 hours	1.01	1 days: +4 hours	0.800
0 days: 0 hours	1.00		
0 days: +1 hours			
0 days: +2 hours			
0 days: +3 hours			
0 days: +4 hours			
0 days: +5 hours			
0 days: +6 hours			
0 days: +7 hours			
0 days: +8 hours			
0 days: +9 hours			
0 days: +10 hours			
0 days: +11 hours			
0 days: +12 hours			
0 days: +13 hours			
0 days: +14 hours			
0 days: +15 hours			
0 days: +16 hours			
0 days: +17 hours			
0 days: +18 hours			
0 days: +19 hours			
0 days: +20 hours			
0 days: +21 hours			
0 days: +22 hours			
0 days: +23 hours			
1 days: 0 hours			

Certificate Calibration Standard

For radioactive solution or suspension:

Product code: RAD-MP-09
 Product name: Calibration sample
 Sample container type: 10 mL Type I vial
 Nominal sample quantity (mL): 11
 Nuclide: Radium-224, 3.8 Days half-life

Technical data
 Batch number: 832100215
 Max radioactivity on transport order: < 10.7 MBq
 Measured activity on 28 Sep 2022: 7.15 MBq/vial at 14:27
 Reference date/time (CET-04 h): 28 Oct 2022/12:00
 Radioactivity at reference date and time: 10.7 MBq/vial

Additional information:

Oncoinvent AS
 Alveya Marvik
 Senior Quality Control Scientist
 Contact information: info@oncoinvent.com Mob: +47 45877226
 Oncoinvent AS - TP: (47) 22 18 35 95 - Gullhaugen 7, 9484 Oslo, Norway - www.oncoinvent.com
oncoinvent@oncoinvent.com - Org. nr. 98876488



Calculation of dose calibrator measurement for dial setting

Measurement	Measured activity (MBq)	Time difference (days: hours)	Decay correction factor	Decay corrected activity (MBq)	Measurement (MBq)	% Difference	Result
Measurement 1	10.5	0 days: 0 hours	1.00	10.5	10.5	0.00	Pass
Measurement 2	10.5	0 days: 0 hours	1.00	10.5	10.5	0.00	Pass
Measurement 3	10.5	0 days: 0 hours	1.00	10.5	10.5	0.00	Pass
Measurement 4	10.5	0 days: 0 hours	1.00	10.5	10.5	0.00	Pass
Measurement 5	10.5	0 days: 0 hours	1.00	10.5	10.5	0.00	Pass
Measurement 6	10.5	0 days: 0 hours	1.00	10.5	10.5	0.00	Pass
Measurement 7	10.5	0 days: 0 hours	1.00	10.5	10.5	0.00	Pass
Measurement 8	10.5	0 days: 0 hours	1.00	10.5	10.5	0.00	Pass
Measurement 9	10.5	0 days: 0 hours	1.00	10.5	10.5	0.00	Pass
Measurement 10	10.5	0 days: 0 hours	1.00	10.5	10.5	0.00	Pass
Measurement 11	10.5	0 days: 0 hours	1.00	10.5	10.5	0.00	Pass
Measurement 12	10.5	0 days: 0 hours	1.00	10.5	10.5	0.00	Pass
Measurement 13	10.5	0 days: 0 hours	1.00	10.5	10.5	0.00	Pass
Measurement 14	10.5	0 days: 0 hours	1.00	10.5	10.5	0.00	Pass
Measurement 15	10.5	0 days: 0 hours	1.00	10.5	10.5	0.00	Pass
Measurement 16	10.5	0 days: 0 hours	1.00	10.5	10.5	0.00	Pass
Measurement 17	10.5	0 days: 0 hours	1.00	10.5	10.5	0.00	Pass
Measurement 18	10.5	0 days: 0 hours	1.00	10.5	10.5	0.00	Pass
Measurement 19	10.5	0 days: 0 hours	1.00	10.5	10.5	0.00	Pass
Measurement 20	10.5	0 days: 0 hours	1.00	10.5	10.5	0.00	Pass
Measurement 21	10.5	0 days: 0 hours	1.00	10.5	10.5	0.00	Pass
Measurement 22	10.5	0 days: 0 hours	1.00	10.5	10.5	0.00	Pass
Measurement 23	10.5	0 days: 0 hours	1.00	10.5	10.5	0.00	Pass
Measurement 24	10.5	0 days: 0 hours	1.00	10.5	10.5	0.00	Pass
Measurement 25	10.5	0 days: 0 hours	1.00	10.5	10.5	0.00	Pass
Measurement 26	10.5	0 days: 0 hours	1.00	10.5	10.5	0.00	Pass
Measurement 27	10.5	0 days: 0 hours	1.00	10.5	10.5	0.00	Pass
Measurement 28	10.5	0 days: 0 hours	1.00	10.5	10.5	0.00	Pass
Measurement 29	10.5	0 days: 0 hours	1.00	10.5	10.5	0.00	Pass
Measurement 30	10.5	0 days: 0 hours	1.00	10.5	10.5	0.00	Pass
Measurement 31	10.5	0 days: 0 hours	1.00	10.5	10.5	0.00	Pass
Measurement 32	10.5	0 days: 0 hours	1.00	10.5	10.5	0.00	Pass
Measurement 33	10.5	0 days: 0 hours	1.00	10.5	10.5	0.00	Pass
Measurement 34	10.5	0 days: 0 hours	1.00	10.5	10.5	0.00	Pass
Measurement 35	10.5	0 days: 0 hours	1.00	10.5	10.5	0.00	Pass
Measurement 36	10.5	0 days: 0 hours	1.00	10.5	10.5	0.00	Pass
Measurement 37	10.5	0 days: 0 hours	1.00	10.5	10.5	0.00	Pass
Measurement 38	10.5	0 days: 0 hours	1.00	10.5	10.5	0.00	Pass
Measurement 39	10.5	0 days: 0 hours	1.00	10.5	10.5	0.00	Pass
Measurement 40	10.5	0 days: 0 hours	1.00	10.5	10.5	0.00	Pass
Measurement 41	10.5	0 days: 0 hours	1.00	10.5	10.5	0.00	Pass
Measurement 42	10.5	0 days: 0 hours	1.00	10.5	10.5	0.00	Pass
Measurement 43	10.5	0 days: 0 hours	1.00	10.5	10.5	0.00	Pass
Measurement 44	10.5	0 days: 0 hours	1.00	10.5	10.5	0.00	Pass
Measurement 45	10.5	0 days: 0 hours	1.00	10.5	10.5	0.00	Pass
Measurement 46	10.5	0 days: 0 hours	1.00	10.5	10.5	0.00	Pass
Measurement 47	10.5	0 days: 0 hours	1.00	10.5	10.5	0.00	Pass
Measurement 48	10.5	0 days: 0 hours	1.00	10.5	10.5	0.00	Pass
Measurement 49	10.5	0 days: 0 hours	1.00	10.5	10.5	0.00	Pass
Measurement 50	10.5	0 days: 0 hours	1.00	10.5	10.5	0.00	Pass

Dose calibrator dial setting for Radspherin®

Background/Calculation	Measurements
Reference standard Batch no. 832090209	Observed dial setting for Radspherin® 6'5
Reference standard activity (MBq) at reference time 10.5	
Reference standard date and time (dd/mm/yyyy hh:mm) 28/09/2022 13:00	
Date and time of dose calibrator dial setting (dd/mm/yyyy hh:mm) 01/10/2022 13:56	Measured activity (MBq) 2'71
Decay corrected activity of reference (MBq) 2'71	
Date and time of Alternative Measurement 1 (dd/mm/yyyy hh:mm) 01/10/2022 13:57	Alternative Measurement 1 (MBq) 2'71
Decay corrected activity of reference (MBq) 2'71	
Date and time of Alternative Measurement 2 (dd/mm/yyyy hh:mm) 01/10/2022 14:00	Alternative Measurement 2 (MBq) 2'71
Decay corrected activity of reference (MBq) 2'71	
Alternative Measurement 2 (dd/mm/yyyy hh:mm) 01/10/2022 14:01	Alternative Measurement 2 (MBq) 2'72
Decay corrected activity of reference (MBq) 2'71	
Dose Calibrator manufacturer: BIODAX	
Dose Calibrator model: ATONLAB 500	
Dose Calibrator serial number: 71435046	
Dose calibration performed by (printed name): ELENA PRIETO	
Email address of contact at calibration site: eprieto@oncoinvent.com	
Dose calibration performed by (signature and date): <i>E. Prieto</i> 01-10-22	



Calibration standard

onco invent
01 Nov 2022

Certificate Calibration Standard

For radioactive solution or suspension:

Product code: RAD-MP-09
Product name: Calibration sample
Sample container type: 10 mL Type I vial
Nominal sample quantity (mL): 1

Nuclide: Radium-224, 3.8 Days half life

Technical data

Batch number: 022190215
Max radioactivity on transport order: < 10.7 MBq
Measured activity on 28 Sep 2022: 7.15 MBq/vial at 14:27
Reference date/time (CET-04 h): 26 Oct 2022/12:00
Radioactivity at reference date and time: 16.7 MBq/vial

Contamination test: Wipe test on glass vial passed: no activity detected on surface of vial
Passed (yes/no): *yes*
Date of test: *26/09/2022*

Additional information:

Oncoinvent AS
Address: Callingsveien 7 0464 Gjøa, Norway
E-mail: oncoinvent@oncoinvent.com
Tel: +47 22 18 33 85
MBA number: 3751-1

Oncoinvent AS
Address: Callingsveien 7 0464 Gjøa, Norway
E-mail: oncoinvent@oncoinvent.com
Tel: +47 22 18 33 85
MBA number: 3751-1

onco invent

Certificate of Analysis

Match number: **03100189**



Item	Test procedure	Specification	Test result
Appearance	434-CP-00075	Clear, colorless solution with white microcrystals	Conforms
Radioactive identity of the 224Ra	434-CP-00070	Identified gamma energy peaks for 224Ra and daughter nuclides	Conforms
Assay radioactivity concentration of reference date and time	434-CP-00288	Target activity Range (MBq/L) (± 10% of target activity)	1138 MBq/mL
Radioactive impurities (226Ra, 228Ac, 228Th, 228Pa, 228Fr, 228Ra, 228Ac, 228Th, 228Pa, 228Fr, 228Ra)	434-CP-01004	10.000 % ±1% 10.000 % ±1% 10.014 % ±1% relative to 224Ra parent dose	TDC % ±1% TDC % ±1% TDC % ±1%
Radioactive purity	434-CP-01004	± 99.9% relative to 224Ra parent dose	TDC %
224Ra adsorbed onto carbon microcrystals	434-CP-00073	± 97%	100%
Bulkiness	434-CP-00074	70 ppm for calculated per g microcrystals	TDC ppm
pH	434-CP-00072	6.5 - 10.0	9.5
Particle size distribution	434-CP-00085	D(0.1) < 2 µm D(0.5) 3 - 20 µm D(0.9) < 120 µm	D(0.1) 4 µm D(0.5) 9 µm D(0.9) 16 µm
Bacterial endotoxins	434-CP-00074	< 8.75 EU/mL	< 8.75 EU/mL
Quality	434-CP-00087	Stable	Conforms
100% quality	434-CP-01005	Stable confirmed	Conforms

Conclusion

1. Fulfilled according to specification: 434-CP-00040, rev. 07
2. Compliance with acceptance criteria: 02 Nov 2022
3. Date test performed/checked: 07 Sep 2021
4. Comments/observations: None

Date and signature:
QC Scientist: *Ingrid Eke*
Sign for Head of Quality Control/Gjøa Elisabeth Høllum

Page 1 of 1





Analytical method: instrument

Library Listing Report 29.06.2022 15:14:22 Page 1

LIBRARY LISTING REPORT



Filename: C:\Canberra\Apex\Root\Default\Library\Radprod 25 apr 2022.M

Nuclide Library Description:

Nuclide Name	Half-Life (Seconds)	Energy (keV)	Uncert. (keV)	Yield (%)	Yield Uncert. (Abs.--)
TL-208	3.1622E+05	277.370	0.000	6.6000	0.0000
		583.190*	0.002	85.0000	0.0000
		880.564	0.005	12.0000	0.1000
Bi-212	3.1622E+05	727.350*	0.040	6.8700	0.1500
Pb-212	3.1622E+05	785.400	0.000	1.1000	0.0000
Bi-214	3.1622E+05	220.433*	0.004	63.6000	1.3000
Po-214	3.1622E+05	300.100	0.000	3.3000	0.0000
Pb-214	3.1622E+05	240.590*	0.006	4.1000	0.0000

* = key line



TOTALS: 4 Nuclides 8 Energy Lines





Analytical method: parameters



- Sample preparation
- Geometry
 - Sample size (volume)
 - Distance from detector
- Counting time
- Sample size (activity -> dead time)
- Interferences (e.g., ^{212}Pb)



Quality control




Analytical method: validation





Analytical method: validation





Parameter	Possible values	Acceptance criteria	Result
Range, kBq/mL (2 days after HPGe measurements ≈ reference date and time)	502-1189	400-1400	314-1581
Range, kBq/mL (on HPGe measurements date)	735-1742	590-2090	461-2316

Quality control



Analytical method: true values – standardization – traceability



Complete chain of production, control and use





onco
invent



Questions?
