

Challenges to the use of the GUM:

- **Issues with the uncertainty evaluation associated with the calibration of indicating instruments**
- **The growing problem with “outlier rejection” in measurements with measurands defined by “extreme values”**

Steve Phillips (NIST)

In collaboration with

Alessandro Balsamo (INRIM)

Michael Krystek (PTB)

Craig Shakarji (NIST)

Industrial problems with the uncertainty associated with the calibration of indicating measuring instruments

- Industrial practice frequently includes incorrect uncertainty contributors
- The problem affects most dimension calibration labs (and others...), worldwide
- The issue involves confusion about the measurand

Survey presented at ISO TC213 WG6 on Digital Caliper calibration uncertainty	Expanded Uncertainty in Verifying Specification		Caliper Accuracy Specification		
	Country	Uncertainty (±)	E_{MPE} (±)	Zone of compliance (Apply ISO 14253-1) (±)	Proof of conformance (Apply ISO 14253-1)
	France	23 μm	60 μm	37 μm	Possible
	UK	16 μm	20 μm	4 μm	Not possible
	Germany	36 μm	30 μm	none	Not possible

Our Message:

- The GUM is “correct” but requires some careful thinking to address this subject; some explicit GUM examples are needed to definitively resolve this issue for practitioners

Calibration of steel artifacts

$0 < L \leq 10 \text{ mm}$



Length
Measurand
defined by
ISO XXX



$L \pm 0.2 @ T=20^\circ\text{C}$


Specification by
artifact designer

Unknown Information
= Value of the
manufacturing Error
= Reference – Design Value

Known Information=
Reference value of the
calibrated Instrument
Ref = indicated value
 $U_{\text{ref}} = 0.030$

Value of Manufacturing Error
= $10.090000 - 10.00000$
= 0.090000

Calibration of a digital caliper for measuring steel artifacts $0 < L \leq 10 \text{ mm}$



Accuracy Measurand
defined by ISO YYY for
one measurement

$MPE = \pm 0.03$

$18^\circ\text{C} \leq T \leq 22^\circ\text{C}$

Specification by
caliper
manufacturer

Known Information =
Reference Value from
the Calibrated Gauge
Ref = 10.000000
 $U_{\text{ref}} = 0.000051$

Unknown Information
= Caliper Error
= Indicated – Ref. Value

Role Reversal

Value of Caliper Error
= $10.020000 - 10.000000$
= 0.020000

Report of Calibration of Steel Artifacts

L	Manufacturing Error	Uncertainty
1.0	0.140	0.030
1.1	-0.030	0.030
...
10.0	0.090	0.030

Quality of Artifact
manufacturing

Quality of Calibration
measurement



The magnitude of the manufacturing errors do not affect the uncertainty of their measurement.

Uncertainty in Manufacturing Error

Source	Std. Unc.	Type	Notes
Ref. Std. (Caliper)	0.015 = MPE/2	B	From calibration report MPE=0.030
Thermal *	0.000005	B	CTE uncertainty
$U_{k=2}(L)$	0.030		

*uncertainty in the CTE value requested by the instrument; Measurement at T = 21 °C

Report of Calibration of Digital Caliper for Measuring Steel Artifacts

L	Caliper Error	Uncertainty
1.0	0.010000	0.000051
1.1	-0.020000	0.000051
...
10.0	0.020000	0.000051

Quality Caliper
manufacturing

Quality of Calibration
measurement



The magnitude of the caliper errors do not affect the uncertainty of their measurement.

Uncertainty in Caliper Error

Source	Std. Unc.	Type	Notes
Ref. Std. (Gauge block)	0.000025	B	From $U/2$ on gauge cal. report
Thermal *	0.000005	B	CTE uncertainty
$U_{k=2}(E)$	0.000051		


*uncertainty in the CTE value of the gauge block; Measurement at T = 21 °C

What should happen next...

Report of Calibration of Steel Artifacts

L	Manufacturing Error	Uncertainty
1.0	0.140	0.030
1.1	-0.030	0.030
...
10.0	0.090	0.030

Use the Cal. Report and perform
“guarded acceptance” with a
guardband = $U = 0.030$




All manufacturing Errors are within the
guarded acceptance zone...

This IS the current industrial practice 😊

Report of Calibration of Digital Caliper for Measuring Steel Artifacts

L	Caliper Error	Uncertainty
1.0	0.010000	0.000051
1.1	-0.020000	0.000051
...
10.0	0.020000	0.000051

Use the Cal. Report and perform
“guarded acceptance” with a
guardband = $U = 0.000051$



All caliper Errors are within the guarded
acceptance zone...

Unfortunately, this is NOT current
industrial practice... ☹


What usually happens next for calibrating indicating instruments...

Report of Calibration of Digital Caliper for Measuring Steel Artifacts

L	Caliper Error	Uncertainty	Uncertainty
1.0	0.010000	0.000051	0.021005
1.1	-0.020000	0.000051	0.021005
...	
10.0	0.020000	0.000051	0.021005

Use the Cal. Report and perform “guarded acceptance” with a guardband = $U = 0.021005$

Source	Std. Unc.	Type	Notes
Ref. Std. (Gauge block)	0.000025	B	From $U/2$ on gauge cal. report
Thermal *	0.000005	B	CTE uncertainty
Repeatability	0.010502	A	from “instrument”
$U_{k=2}(L)$	0.021005		



Many caliper Errors are OUTSIDE the guarded acceptance zone...

The Mistake is thinking that the caliper is the “instrument” – it is NOT – it is the unknown quantity under test. The “instrument” is the known quantity – the calibrated gauge block, and its small repeatability is already included via the thermal uncertainty source.

Proper evaluation of Repeatability will reduce the uncertainty associated with verification to specifications to $< 0.1 \mu\text{m}$

Survey presented at ISO TC213 WG6 on digital caliper calibration uncertainty	Country	Uncertainty (\pm)	E_{MPE} (\pm)	Zone of compliance (Apply ISO 14253-1) (\pm)	Proof of conformance (Apply ISO 14253-1)
	France	23 μm	60 μm	37 μm	Possible
	UK	16 μm	20 μm	4 μm	Not possible
	Germany	36 μm	30 μm	none	Not possible

The caliper repeatability is included in the MPE, and it is the MPE that characterizes the uncertainty of the caliper when evaluating the uncertainty of a subsequent measurement made using the caliper

Helpful Hints for Determining Uncertainty Sources for Indicating Instrument Calibrations

- Think about:
 - the “known information” providing the calibration
 - the “unknown information” under calibration
- The term “instrument” needs to be clearly associated with the Known Information

From This Perspective...

- Uncertainty sources arise from:
 - Imperfect information about the known quantity, i.e., the reference value (including its repeatability!)
 - Imperfect realizations of required measurement conditions that **DO** create dispersion in the Errors.



A Caliper is the indicating instrument under calibration

$MPE = \pm 0.03$ when $18^{\circ}\text{C} \leq T \leq 22^{\circ}\text{C}$

Measurement at 21°C

Imperfect CTE Information: $11 < \text{CTE} < 12$


		Uncertainty Source?	
		YES	NO
1.	Caliper <u>Requires</u> the user entered CTE value for normal operation: ✓		
2.	Caliper <u>Does Not Require</u> the CTE value for normal operation:		✓
3.	Caliper <u>Requires Selection</u> of CTE via a pull down menu listing materials: “glass”, “steel”, “aluminum” & We select “steel”:		✓
4.	Caliper <u>Requires Selection</u> of CTE via a pull down menu listing selected values: “10.5”, “11.5”, “12.5” & We select “11.5”:		✓
5.	Caliper <u>Requires Selection</u> of CTE via a pull down menu listing selected values: “10.4”, “11.1”, “11.8”, “12.5” & We select “11.8”: ✓		

Summary of Measurement Uncertainty Associated with the Calibration of Indicating Instruments

- Widespread problem in industry & calibration labs
- The GUM provides a consistent and unified approach to the calibration of both artifacts and instruments but careful consideration is needed to successfully address each case
- We believe the problem has a well-defined solution:
GUM “instrument” \equiv system providing the Known Information
- Industrial practitioners would greatly benefit from explicit examples in the GUM




The growing problem in with “outlier rejection” in measurements with measurands defined by “extreme values”

- In Dimensional Metrology (DM) the measurands of geometrical features, e.g., size, location, orientation and form, are determined entirely by the “extreme” points on the feature surface






- For example: The Flatness of a plane is defined the distance between two parallel planes.

(Only a few points on the physical surface determines the flatness value)



- In the past 20 years DM measurements have increased point coordinate collection from $10^2 \rightarrow 10^7$ points.
- Outlier rejection is now a significant problem because:
 - the “real” extreme points that determine the true value of the measurand may also be rejected by the outlier filter...
 - Averaging repeated measurement results converges on the magnitude of the mean erroneous outlier value...



The GUM Does Comment on Blunders & Mistakes... But Not on Outliers...

- No consistent definition of outliers...
- Outlier filters are widely used but not reported...
- Changing filters on the same raw data set can change the measured value (e.g., by 10 U), but no uncertainty source for this in the U budget
- Some argue that outliers are mistakes and hence do not need to be addressed by uncertainty
- The GUM is silent on this topic...
 - Even a short paragraph about this issue in the GUM would be valuable!

Closing Thoughts

- Uncertainty of Calibrating Indicating instruments
 - A wide spread problem
 - Has a well-defined solution: Examples in the GUM
- Outlier rejection of large data sets
 - A wide spread problem
 - Some general guidance in the GUM would be valuable

Thank You For the Opportunity to
Discuss these Issues