



Bureau International des Poids et Mesures

JCTLM SYMPOISUM 2007



Joint Committee for Traceability in Laboratory Medicine

**A Database of Higher Order Reference Materials
and Reference Measurement Procedures**

R.I. Wielgosz, BIPM

www.bipm.org

Joint Committee for **Traceability** in Laboratory Medicine

- What is it?
- Why is it important?
- How is it implemented?
- How does the JCTLM help?

What Is Metrological Traceability?

- Definition VIM:1993, 6.10

Metrological traceability - property of the result of a measurement or the value of a standard whereby it can be related to **stated references**, usually national or international standards, **through an unbroken chain of comparisons** all having stated **uncertainties**

NOTES

- 1 The concept is often expressed by the adjective **traceable**.
- 2 The unbroken chain of comparisons is called a **traceability chain**.

Uncertainty Definition

VIM 1993, 3.9

- **uncertainty of measurement** - parameter, associated with the result of a measurement, that characterizes the dispersion of the values that could reasonably be attributed to the measurand

Glucose in Blood, Serum, Urine, CSF
SI-Unit: mmol/l

*Definition
of the measurand*

Section 1 –External to manufacturer, credentialing of the Certified Reference Material

SRM917b

NIST certification of SRM917b (purity)

Primary calibrator

SRM917b – weighed amount

Weighing procedure

Primary reference measurement procedure

Secondary calibrator

Human Patient Specimens, e.g. Blood, Serum, Urine, CSF

Higher Order Reference Procedure – e.g. Isotope Dilution - Mass Spectrometry or Procedure of Similar Trueness and Precision

Secondary reference measurement procedure

Section 2 –Internal to manufacturer, value assignment

Manufacturer's working calibrator

Manufacturer's Master Calibrator, Master Lot of Product Calibrator

Reference Procedure traceable to higher order reference procedure - e.g. Hexokinase/glucose-6-phosphate Dehydrogenase Procedure

Manufacturer's selected measurement procedure

Product Calibrator

New Lot Commercial Product Calibrator

Procedure applying same chemistry and equipment as routine procedure, but more precisely controlled conditions and more replicates to reduce uncertainty

Manufacturer's standing measurement procedure

Section 3 –External to manufacturer, End user's results are Traceable to Certified Reference Material and the Reference System

Commercially available system including product reagent and calibrator lots

End user's routine measurement procedure

Routine Sample – Human Patient Specimens, e.g. Blood, Serum, Urine or CSF

RESULT
Glucose in mmol/l

Traceability Chain for the Measurement of Glucose in Body Fluids

The Importance of Metrological Traceability

- Why has traceability become important
 - Clinical Considerations
 - Legislative and Standardization Efforts
- Ensuring Metrological Traceability
 - The Traceability Chain
 - Commutability
 - Uncertainty

Why Traceability

- Evidence Based Medicine is becoming more and more prominent in the practice of clinical laboratory science
 - Glucose
 - PSA
 - HbA1c
 - Creatinine (GFR calculations)
- All of these analytes have specific cutoffs that are independent of the assay used.
- To correctly utilize these cutoffs, the assays for the analyte in question must be comparable.

Comparability

- For the above types of clinical guidelines to be effective, the results from clinical laboratories must be comparable
- To be comparable they must to “traceable”
 - i.e. a traceability chain to the same materials and methods, commutable with patient sera and an uncertainty within clinical guidelines

Regulatory Requirements

- **IVD Directive (IVD Manufacturers)**
“The traceability of values assigned to calibrators and/or control materials must be assured through available reference measurement procedures and/or available reference materials of a higher order”.
- **ISO 18159 (Clinical Laboratories)**
“A program for calibration of measuring systems and verification of trueness shall be designed and performed so as to ensure that measurements are traceable to the SI units or by reference to a natural constant”

“The laboratory shall determine the uncertainty of its measurements, where relevant and possible”.
- **ISO 17025 has similar requirements**



What has JCTLM delivered?

A Quality assured database of:

- a) Higher Order Reference Materials**
- b) Reference Measurement Procedures**
- c) Laboratory Reference Measurement Services (2007)**

<http://www.bipm.org/jctlm/>

For use by (primarily)

- a) IVD industry**
- b) Regulators**

Glucose in Blood, Serum, Urine, CSF
SI-Unit: mmol/l

**Combined
standard
uncertainty (%)**

Section 1 –External to manufacturer, credentialing of the Certified Reference Material

SRM917b

NIST certification of SRM917b (purity)

0.1%

Primary calibrator

SRM917b – weighed amount

Weighing procedure

0.87%

Secondary calibrator

Human Patient Specimens, e.g. Blood, Serum, Urine, CSF

Higher Order Reference Procedure – e.g. Isotope Dilution - Mass Spectrometry or Procedure of Similar Trueness and Precision

Section 2 –Internal to manufacturer, value assignment

Manufacturer's working calibrator

Manufacturer's Master Calibrator, Master Lot of Product Calibrator

Reference Procedure traceable to higher order reference procedure - e.g. Hexokinase/glucose-6-phosphate Dehydrogenase Procedure

1.21%

Product Calibrator

New Lot Commercial Product Calibrator

Procedure applying same chemistry and equipment as routine procedure, but more precisely controlled conditions and more replicates to reduce uncertainty


1.49%

Section 3 –External to manufacturer, End user's results are Traceable to Certified Reference Material and the Reference System

Commercially available system including product reagent and calibrator lots

Routine Sample – Human Patient Specimens, e.g. Blood, Serum, Urine or CSF

RESULT
Glucose in mmol/l



Glucose in Blood, Serum, Urine, CSF
SI-Unit: mmol/l

**Combined
standard
uncertainty (%)**

Section 1 –External to
manufacturer, credentialing of the
Certified Reference Material

JCTLM ACTIVITIES

ISO 15193, ISO 15194, ISO 15195

Characterization of SRM917b

0.1%

Weighting procedure

Higher Order Reference Procedure –
e.g. Isotope Dilution - Mass
Spectrometry or Procedure of Similar
Trueness and Precision

0.87%

Secondary calibrator

Human Patient Specimens,
e.g. Blood, Serum, Urine,
CSF

Section 2 –Internal to
manufacturer, value assignment

Reference Procedure traceable to
higher order reference procedure -
e.g. Hexokinase/glucose-6-
phosphate Dehydrogenase
Procedure

1.21%

*Manufacturer's
working calibrator*

Manufacturer's Master
Calibrator, Master Lot of
Product Calibrator

Procedure applying same chemistry
and equipment as routine procedure,
but more precisely controlled
conditions and more replicates to
reduce uncertainty

1.49%

Product Calibrator

New Lot Commercial
Product Calibrator

Section 3 –External to
manufacturer, End user's results are
Traceable to Certified Reference Material
and the Reference System

Commercially available system
including product reagent and
calibrator lots

Routine Sample – Human Patient Specimens,
e.g. Blood, Serum, Urine or CSF

ISO 15189

RESULT
Glucose in mmol/l



Bureau International des Poids et Mesures

Database of higher-order reference materials and reference measurement methods/procedures



JCTLM Database
Laboratory medicine and *in vitro* diagnostics

> You are here : JCTLM-DB home

T+ T T-

JCTLM database: Laboratory medicine and *in vitro* diagnostics

[Version française](#)

JCTLM-DB

- Advanced search
- Preamble for JCTLM Lists
- Reference materials no longer listed
- Contact us

JCTLM

- JCTLM
- JCTLM Working Group 1
- JCTLM Working Group 2



Search by analyte

Type an analyte name in part or full, e.g. cholesterol

Please select your requirement :

- Higher-order reference materials
- Reference measurement methods/procedures

Search

Advanced search

JCTLM Database

In response to the requirements of the European Directive on *in vitro* diagnostic devices, the [Joint Committee for Traceability in Laboratory Medicine](#) (JCTLM) publishes lists of available higher-order reference materials and reference measurement methods/procedures for IVD devices.

Higher-order reference materials and reference measurement methods or procedures published in the JCTLM database are categorized in two lists:

- List I: Certified reference materials and reference measurement methods/procedures for well-defined chemical entities or internationally recognized reference method-defined measurands. Reference materials and measurement methods/procedures included in this category are those that provide values that are traceable to the SI units; e.g., electrolytes, enzymes, drugs, metabolites and substrates, non-peptide hormones, and some proteins.



Operational Structure

- **Executive Committee**
- **Working Groups**
 1. **Reference Materials and Reference Methods**

Chairs: H. Schimmel (IRMM), W. May (NIST)

Compilation of existing RMPs and CRMs (Lists)
 2. **Reference Laboratories - Networks**

Chairs: L. Siekmann, L. Thienpont (IFCC)

Guidelines for reference laboratories

Identification of networks

Relevant ISO Standards for higher order RMs and RMPs

ISO 17511 In vitro diagnostic medical devices - Measurement of quantities in biological samples - Metrological traceability of values assigned to calibrators and control materials

ISO 18153 Metrological traceability of values for catalytic concentration of enzymes assigned to calibrators and control materials

ISO 15193 Presentation of reference measurement procedures

ISO 15194 Description of reference materials

ISO 15195 Reference Measurement Laboratories

JCTLM WG1 Review Process

INTERNATIONAL
STANDARD

ISO
15194

First edition
2002-10-01

***In vitro* diagnostic medical devices —
Measurement of quantities in samples of
biological origin — Description of reference
materials**

*Dispositifs médicaux de diagnostic in vitro — Mesure des grandeurs dans
les échantillons d'origine biologique — Description des matériaux de
référence*



Reference number
ISO 15194:2002(E)

© ISO 2002

JCTLM WG1 Review Process (ISO 15194)

5 Description of a reference material

5.1 Elements of a description

The description of a reference material of higher metrological order shall comprise at least the elements listed as mandatory (M) in Table 1.

NOTE: The order of the elements listed in Table 1 may be changed and additional elements, such as an abstract, may be added as appropriate.

Table 1: Main elements (clauses) of a report describing a reference material of higher metrological order

| Element | Type ¹⁾ | | Subclause in this European Standard |
|--|--------------------|---|-------------------------------------|
| | M | O | |
| Title page | I | | |
| Contents list | | I | |
| Foreword | I | | |
| Warning and safety precautions | N | | 5.2 |
| Introduction | | I | 5.3 |
| Title of report | N | | |
| Scope | N | | 5.4 |
| Definitions | | N | |
| Symbols and abbreviations | | N | |
| Terminology | | N | 5.5 |
| Justification for choice of reference material | I | | 5.6 |
| General characteristics | I | | 5.7 |
| Specific characteristics | I | | 5.8 |
| Validation | I | | 5.9 |
| Intended function | I | | 5.10 |
| Instructions for use | I | | 5.11 |
| Supplier | I | | 5.12 |
| Bibliography | | I | 5.13 |
| Annexes | | I | 5.14 |
| Dates | I | | 5.15 |
| ¹⁾ Symbols for type of element in a European Standard: M mandatory, O optional; I informative, N normative. | | | |

JCTLM-WG1 Quality Manual:
Reference Materials and Reference Procedures



| | | | | |
|---|---|--|-----|------------|
| ↙ | <u>Preamble</u> | | PDF | 2006/01/31 |
| ↙ | <u>WG1-P-00</u> | Quality Policy | PDF | 2006/01/30 |
| ↙ | <u>WG1-P-01</u> | Outline of JCTLM procedures | PDF | 2006/01/31 |
| ↙ | <u>WG1-P-02</u> | Reference material and procedure nomination requests | PDF | 2006/01/31 |
| ↙ | <u>WG1-P-02-F-01 & F-02</u> | Reference material and procedure nomination templates | ✘ | 2006/01/30 |
| ↙ | <u>WG1-P-02-F-01 EXAMPLE</u> | Reference material nomination example | ✘ | 2006/01/30 |
| ↙ | <u>WG1-P-02-F-02 EXAMPLE</u> | Reference procedure nomination example | ✘ | 2006/01/30 |
| ↙ | <u>WG1-P-02-I-01</u> | Instructions for completing nomination templates | PDF | 2006/01/31 |
| ↙ | <u>WG1-P-03</u> | Review and approval of nominations | PDF | 2006/01/30 |
| ↙ | <u>WG1-P-03-F-03</u> | Review report form | DOC | 2006/01/30 |
| ↙ | <u>WG1-P-04A</u> | Multiple CRM comparison process | PDF | 2006/01/30 |
| ↙ | <u>WG1-P-04B</u> | Multiple reference method/procedure comparison process | PDF | 2006/01/30 |
| ↙ | <u>WG1-P-05</u> | Consensus review and communication of recommendations | PDF | 2006/01/30 |
| ↙ | <u>WG1-P-06</u> | Membership of JCTLM WG1 Review Teams | PDF | 2006/01/31 |
| ↙ | <u>WG1-P-06-F-01</u> | Review Team membership form | DOC | 2006/01/30 |
| ↙ | <u>WG1-P-07</u> | Process for changing the WG1 Quality System procedures | PDF | 2006/01/31 |
| ↙ | <u>WG1-P-07-F-01</u> | Procedure change request form | DOC | 2006/01/30 |
| ↙ | <u>WG1-P-08</u> | Process for changing WG Review Teams | PDF | 2006/01/30 |
| ↙ | <u>WG1-P-09</u> | Appeals process | PDF | 2006/01/30 |

LIST I

published initially on 01 April 2004

Certified Reference Materials and Reference Measurement Procedures for well-defined chemical entities or internationally recognized reference method-defined measurands, such as enzymes. Reference Materials included in this category are those that are traceable to the SI units. [***Electrolytes, Drugs, Metabolites and Substrates, Non-Peptide Hormones, Enzymes and some Proteins***]

approximately **123** Reference Measurement Procedure entries for **75** different health status markers

approximately **211** Reference Material entries for **128** measurands

<http://www.bipm.org/en/committees/jc/jctlm/jctlm-db/>

JCTLM LIST II

Reference Materials that are value-assigned using an internationally agreed upon protocol e.g., reference materials for Blood Typing, Coagulation Factors, Microbial Serology, Nucleic Acids, and some Proteins. The values of the measurands in the reference materials on this List are not SI-traceable and/or no internationally-recognized reference measurement procedures exist.

Initially published in January, 2005, now includes

- 12 CRMs for Coagulation Factors
- 7 CRMs for Proteins

Hereafter, Lists I and II to be updated in April of each year

Summary of Cycles I & II Nominations and Approvals

| Category | Number of Nominations Submitted | | Number Recommended for Publication | |
|----------------------------|---------------------------------|----------------------------------|------------------------------------|----------------------------------|
| | Reference Materials | Reference Measurement Procedures | Reference Materials | Reference Measurement Procedures |
| Blood Gases | 1 | 1 | 0 | 0 |
| Drugs | 84 | 5 | 23 | 3 |
| Electrolytes | 70 | 23 | 29 | 23 |
| Enzymes | 20 | 7 | 11 | 6 |
| Metabolites and Substrates | 69 | 44 | 39 | 35 |
| Non-Electrolyte Metals | 43 | 50 | 30 | 15 |
| Non-Peptide Hormones | 15 | 26 | 14 | 22 |
| Nucleic Acids | 5 | 0 | 0 | 0 |
| Vitamins | 8 | 2 | 7 | 0 |
| Proteins | 114 | 20 | 43 | 19 |
| Blood Groupings | 3 | 0 | 0 | 0 |
| Coagulation Factor | 34 | 0 | 12 | 0 |
| Microbial Serology | 10 | 8 | 0 | 0 |
| Other | 6 | 2 | 3 | 0 |
| TOTAL Number | 482 | 188 | 211 | 123 |

All Listed Reference Materials will be assessed for comparability by a reference measurement procedure under repeatability conditions:

- to assess veracity of the Normative Standards-Based Review Process
- To establish bias that could be introduced by randomly selecting any material from the List

Example: Potassium in Human Serum CRMs on provisional JCTLM List 1 were assessed for comparability by a single laboratory (NIST) using a reference measurement procedure under repeatability conditions.

Fig 1: Ratio Display




Fig 2: Scattergram Display






Fig 1: The vertical axis reports the ratio between the measured and certified values of each CRM, X/C_i . The dark vertical lines represent the approximate 95% CI about the ratios. The light horizontal lines represent the certified 95% CIs. The dotted line represents the expected ratio for the suite of all materials given the observed identity between the measured and certified values. (CI = Confidence Interval)

Fig 2: The data demonstrate that these CRMs are comparable over a wide concentration range. The horizontal axis reports the certified values, C_i ; the vertical reports the average measured values, X_i . Each level of each CRM is displayed as approximate 95% CIs along both axes. The intersection of these intervals is bounded by an open circle to aid visual inspection.

Comparability of Cholesterol in Serum CRMs on JCTLM LIST



⇒ CRM comparability independent of analyte level

The measured/certified ratios for this set of CRMs are:


- ~ normally distributed
- with a standard deviation of ~0.7%

JCTLM Working Group II - Reference Measurement Services

Assessed on basis of:

- **Metrological level of the reference procedures used (compliance with JCTLM WG1 RMPs)**
- **Accreditation to ISO 17025 and ISO 15195 as a calibration laboratory and/or CIPM-MRA process**
- **Ongoing performance in inter-laboratory comparisons**

SI traceability - Primary Calibrators (ISO 17511)



JCTLM Working Group II - Reference Measurement Services for Laboratory Medicine Cycle I Call for nominations (Closing date 30 April 2006)

JCTLM-WG2 Procedure Manual:
Reference Measurement Services



| | | Title | File type | Last update |
|---|--------------------------------------|--|-----------|-------------|
| ↓ | <u>WG2-P-00</u> | Outline of calibration hierarchy in laboratory medicine | PDF | 2006/01/30 |
| ↓ | <u>WG2-P-01</u> | Overview of JCTLM WG2 procedures | PDF | 2006/01/30 |
| ↓ | <u>WG2-P-02</u> | Nomination process for reference measurement laboratory services | PDF | 2006/01/31 |
| ↓ | <u>WG1-P-02-F-01</u> | Template for nomination of reference laboratory measurement services | | 2006/01/30 |
| ↓ | <u>WG2-P-03A</u> | Review of NMI measurement services | PDF | 2006/01/30 |
| ↓ | <u>WG2-P-03B</u> | Review of measurement services from accredited laboratories | PDF | 2006/01/30 |
| ↓ | <u>WG2-P-03C</u> | Review of measurement services from laboratories preparing for accreditation | PDF | 2006/01/30 |
| ↓ | <u>WG2-P-04</u> | Listing reference measurement laboratory services | PDF | 2006/01/30 |
| ↓ | <u>WG2-P-05</u> | De-listing reference measurement laboratory services | PDF | 2006/01/30 |



Reference Laboratories in Laboratory Medicine

Registration Form for Ring Trial RELA 01/2004

Reference Laboratory Address:

| | |
|-----------------------------|------------|
| Organisation: | Laboratory |
| Name of person responsible: | Name |
| Street: | |
| City: | |
| Post Code: | |
| State: | |
| Country: | |
| Phone: | |
| Fax: | |
| e-mail: | |

Please return by August 15, 2004 to:

Dr. R. Kruse,
Dr. W. Geilenkeuser
DGKL
Im Mühlenbach 52 a
D-53127 Bonn - Germany

Fax: +49-228-211529
E-mail: info@dgkl-rfb.de


Please switch to next table (Measurands) !

| | A | B | C |
|----|-------------------------------------|---|-----------------|
| 11 | | | |
| 12 | Measurand | Participation please indicate "X" | Number of Vials |
| 13 | | | |
| 14 | | | |
| 15 | | | |
| 16 | METABOLITES & SUBSTRATES | | |
| 17 | Total Cholesterol | | |
| 18 | Total Glycerol | | |
| 19 | Creatinine | x | 2 x 5 |
| 20 | Uric Acid | | |
| 21 | Urea | | |
| 22 | Glucose | | |
| 23 | Total Bilirubine | | |
| 24 | | | |
| 25 | ELECTROLYTES | | |
| 26 | Sodium | | |
| 27 | Potassium | x | 2 x 5 |
| 28 | Chloride | | |
| 29 | Calcium | | |
| 30 | Lithium | | |
| 31 | Magnesium | | |
| 32 | | | |
| 33 | ENZYMES | | |
| 34 | ALT | | |
| 35 | AST | | |
| 36 | CK | | |
| 37 | LDH | | |
| 38 | GGT | | |
| 39 | Amylase | | |
| 40 | | | |
| 41 | Total Protein | | |
| 42 | | | |
| 43 | HORMONES | | |
| 44 | Aldosterone | | |
| 45 | Cortisol | x | 2 x 5 |
| 46 | Progesterone | | |
| 47 | Testosterone | | |
| 48 | Estradiol-17β | | |
| 49 | Estriol (non-conjugates) | | |
| 50 | Total Thyroxine (TT4) | | |
| 51 | Total Tri-iodothyronine (TT3) | | |
| 52 | 17-Hydroxyprogesterone | | |
| 53 | | | |
| 54 | THERAP.DRUGS | | |
| 55 | Digoxin | x | 2 x 5 |
| 56 | Digitoxin | | |
| 57 | Theophyllin | | |

IFCC EQAS, Inter-laboratory comparisons for Reference Measurement Laboratories

RELA 1/2003

Total cholesterol [mg/dl]




| Lab | A | p.e.u. | B | p.e.u. | method |
|-----|---------|--------|---------|--------|------------------------------|
| 01 | 230,888 | 3,572 | 166,409 | 3,175 | ID-MS |
| 05 | 234,43 | 5,523 | 168,31 | 4,641 | ID-MS |
| 08 | 229,227 | 4,028 | 165,907 | 2,262 | ID-MS |
| 11 | 230,2 | 4,613 | 167,4 | 2,369 | ID-MS |
| 12 | 237,8 | 1,558 | 171,9 | 0,890 | spectrometry (Abell-Kendall) |
| 16 | 237,490 | 2,764 | 171,389 | 1,962 | spectrometry (Abell-Kendall) |
| 18 | 233,5 | 3,08 | 169,3 | 2,42 | HPLC |
| 19 | 238,1 | 0,860 | 170,8 | 1,610 | spectrometry (Abell-Kendall) |
| 27 | 230,888 | 3,919 | 166,023 | 3,105 | ID-MS |

<http://www.dgkl-rfb.de:81>


JCTLM Database: Summary of entries by Region of Origin

Certified Reference Materials



Entries = 202

Reference Measurement Services



Entries = 98

Establishing and validating a traceable measurement system

SI traceability Example

Glucose in Blood, Serum, Urine, CSF
SI-Unit: mmol/l

**Combined
standard
uncertainty (%)**

Section 1 –External to manufacturer, credentialing of the Certified Reference Material

SRM917b

NIST certification of SRM917b (purity)

0.1%

Primary calibrator

SRM917b – weighed amount

Weighing procedure

0.87%

Secondary calibrator

Human Patient Specimens, e.g. Blood, Serum, Urine, CSF

Higher Order Reference Procedure – e.g. Isotope Dilution - Mass Spectrometry or Procedure of Similar Trueness and Precision

Section 2 –Internal to manufacturer, value assignment

Manufacturer's working calibrator

Manufacturer's Master Calibrator, Master Lot of Product Calibrator

Reference Procedure traceable to higher order reference procedure - e.g. Hexokinase/glucose-6-phosphate Dehydrogenase Procedure

1.21%

Product Calibrator

New Lot Commercial Product Calibrator

Procedure applying same chemistry and equipment as routine procedure, but more precisely controlled conditions and more replicates to reduce uncertainty

1.49%


Section 3 –External to manufacturer, End user's results are Traceable to Certified Reference Material and the Reference System

Commercially available system including product reagent and calibrator lots

Routine Sample – Human Patient Specimens, e.g. Blood, Serum, Urine or CSF

RESULT
Glucose in mmol/l





JCTLM Database: Glucose Reference Materials



Bureau International des Poids et Mesures

Database of higher-order reference materials
and reference measurement methods/procedures



JCTLM Database
Laboratory medicine and *in vitro* diagnostics

> You are here : [JCTLM-DB home](#) > [Search form](#) > Reference materials



Results of the search for higher-order reference materials

↳ JCTLM-DB

- ↳ [Advanced search](#)
- ↳ [Preamble for JCTLM Lists](#)
- ↳ [Reference materials no longer listed](#)
- ↳ [Contact us](#)

↳ JCTLM

- ↳ [JCTLM](#)
- ↳ [JCTLM Working Group 1](#)
- ↳ [JCTLM Working Group 2](#)

→ **Your search criteria:** Higher-order reference materials; Analyte: glucose; Analyte category: -; Matrix category: -

↳ Results of the search

Your search criteria produced 2 results.

Select one or several higher-order reference material summary descriptions amongst the following list and click on 'View' to access more information.

↳ [Select all items from the list](#)

Sort by : Analyte Matrix/Material Organization

| Select | Analyte | Analyte category | Matrix/Material | Organization |
|--------------------------|---------|----------------------------|------------------------------|--------------|
| <input type="checkbox"/> | glucose | metabolites and substrates | glucose crystalline material | NIST |
| <input type="checkbox"/> | glucose | metabolites and substrates | human serum | NIST |

↳ [Deselect all items from the list](#)

[View](#) →

↳ [Modify your selection](#)

JCTLM Database: Glucose Reference Materials

Results of the search

glucose in glucose crystalline material

National Institute of Standards and Technology (NIST), United States

Phone: +1 301 975 6776

Email: srminfo@nist.gov

Fax: +1 301 948 3730

Web: <http://www.nist.gov/srm>

| | |
|---|--|
| Name of the reference material | SRM 917b, D-glucose (dextrose) |
| Quantity | Mass fraction |
| Analyte certified/assigned value | 99.7 % |
| Expanded uncertainty (level of confidence 95%) | 0.2 % |
| Reference(s) on commutability | Not applicable: a high-purity material used as a primary calibrator for higher order reference methods |
| Traceability | SI |
| CRM listing | List I |

glucose in human serum

National Institute of Standards and Technology (NIST), United States

Phone: +1 301 975 6776

Email: srminfo@nist.gov

Fax: +1 301 948 3730

Web: <http://www.nist.gov/srm>

| | |
|---|---|
| Name of the reference material | SRM 965a, glucose in frozen human serum |
| Quantity | Amount-of-substance concentration |
| Analyte certified/assigned value | 1.918 mmol/l to 16.24 mmol/l |
| Expanded uncertainty (level of confidence 95%) | 0.02 mmol/l to 0.19 mmol/l |
| Other relevant publication(s) | Method used for certification: <i>Biomed. Mass Spectrom.</i> , 1982, 9 , 395-405 |
| Traceability | SI |
| CRM listing | List I |

Glucose in Blood, Serum, Urine, CSF
SI-Unit: mmol/l

**Combined
standard
uncertainty (%)**

Section 1 –External to manufacturer, credentialing of the Certified Reference Material

SRM917b

NIST certification of SRM917b (purity)

0.1%

Primary calibrator

SRM917b – weighed amount

Weighing procedure

0.87%

Secondary calibrator

Human Patient Specimens, e.g. Blood, Serum, Urine, CSF

Higher Order Reference Procedure – e.g. Isotope Dilution - Mass Spectrometry or Procedure of Similar Trueness and Precision

1.21%

Section 2 –Internal to manufacturer, value assignment

Manufacturer's working calibrator

Manufacturer's Master Calibrator, Master Lot of Product Calibrator

Reference Procedure traceable to higher order reference procedure - e.g. Hexokinase/glucose-6-phosphate Dehydrogenase Procedure

1.49%

Product Calibrator

New Lot Commercial Product Calibrator


Procedure applying same chemistry and equipment as routine procedure, but more precisely controlled conditions and more replicates to reduce uncertainty

Section 3 –External to manufacturer, End user's results are Traceable to Certified Reference Material and the Reference System

Commercially available system including product reagent and calibrator lots


Routine Sample – Human Patient Specimens, e.g. Blood, Serum, Urine or CSF

RESULT
Glucose in mmol/l



Comparison of NMI capabilities for Glucose in Serum Measurements


CCQM-K11



IFCC EQUAS results for two different reference methods for glucose in serum

RELA 2004

Glucose [mg/dl]



B

| Lab | A | p.e.u. | B | p.e.u. | method |
|-----|---------|--------|---------|--------|--------------|
| 03 | 108,612 | 1,081 | 217,297 | 2,162 | ID-MS |
| 05 | 109,36 | 1,86 | 218,41 | 3,71 | ID-MS |
| 24 | 109,2 | 5,59 | 215,4 | 10,75 | spectrometry |
| 27 | 110,0 | 1,45 | 219,0 | 1,97 | ID-MS |

JCTLM Database: Reference Methods for Glucose



Bureau International des Poids et Mesures

Database of higher-order reference materials
and reference measurement methods/procedures



JCTLM Database
Laboratory medicine and *in vitro* diagnostics

> You are here : [JCTLM-DB home](#) > [Search form](#) > Reference measurement methods/procedures



Result of the search for reference measurement methods/procedures

↳ JCTLM-DB

- ↳ [Advanced search](#)
- ↳ [Preamble for JCTLM Lists](#)
- ↳ [Reference materials no longer listed](#)
- ↳ [Contact us](#)

↳ JCTLM

- ↳ [JCTLM](#)
- ↳ [JCTLM Working Group 1](#)
- ↳ [JCTLM Working Group 2](#)

➔ **Your search criteria:** Reference measurement methods/procedures; Analyte: glucose; Analyte category: -; Matrix category: -

↳ Results of the search

Your search criteria produced 2 results.

For more information on a reference measurement method/procedure for a given Analyte/Matrix (or Material)/Measurement principle (or technique) combination, select one or more of the options below.

↳ [Select all items from the list](#)

Sort by : Analyte Measurement principle/technique Matrix/Material

| Select | Analyte | Measurement principle/technique | Matrix/Material |
|--------------------------|---------|------------------------------------|-----------------|
| <input type="checkbox"/> | glucose | Isotope dilution mass spectrometry | blood serum |
| <input type="checkbox"/> | glucose | Spectrophotometry | blood serum |

↳ [Deselect all items from the list](#)

[View](#) ➔

↳ [Modify your selection](#)

JCTLM Database: Reference Methods for Glucose

↳ Results of the search

| Isotope dilution mass spectrometry methods for glucose in blood serum | |
|--|--|
| ▶ NIST definitive method for serum glucose | |
| Applicable matrice(s) | lyophilized, fresh, or frozen human serum |
| Full description of technique(s) | ID/GC/MS |
| Quantity | Amount-of-substance concentration |
| Applicable range | 2 mmol/l to 20 mmol/l |
| Expected uncertainty (level of confidence 95%) | 0.5 % to 1.5 % |
| Reference(s) | <i>Biomed. Mass. Spectrom.</i> , 1982, 9 , 395-405 |
| Comparability assessment study(ies) | <i>Metrologia</i> , 2003, 40 , <i>Tech. Suppl.</i> , 08003 |
| Comment(s) | The expanded uncertainty is relative |
| JCTLM DB identification number | NRMeth 80 |
| ▶ University of Ghent reference method for glucose | |
| Applicable matrice(s) | lyophilized, fresh, or frozen human serum |
| Full description of technique(s) | ID/GC/MS |
| Quantity | Amount-of-substance concentration |
| Applicable range | 1 mmol/l to 20 mmol/l |
| Expected uncertainty (level of confidence 95%) | 1 % to 2 % |
| Reference(s) | <i>Clin. Chem.</i> , 1993, 39 , 1001-1006 <i>Clin. Chem.</i> , 1993, 39 , 993-1000 <i>Eur. J. Clin. Chem. Clin. Biochem.</i> , 1996, 34 , 853-860 |
| Comparability assessment study(ies) | EUROMET 563 |
| Comment(s) | The expanded uncertainty is relative |
| JCTLM DB identification number | NRMeth 4 |

JCTLM Database: Reference Methods for Glucose

↘ Results of the search

Spectrophotometry method for glucose in blood serum

▸ CDC Hexokinase reference method for glucose

| | |
|---|---|
| Applicable matrixe(s) | human serum |
| Full description of technique(s) | Spectrophotometry |
| Quantity | Amount-of-substance concentration |
| Applicable range | 2.78 mmol/l to 22.2 mmol/l |
| Expected uncertainty (level of confidence 95%) | 0.39 % |
| Reference(s) | Neese, JW, et al., HEW Pub No. (CDC) 77-8330, HEW, USPHS, CDC, 1976 |
| Comment(s) | The expanded uncertainty is relative |
| JCTLM DB identification number | NRMeth 34 |

Glucose in Blood, Serum, Urine, CSF
SI-Unit: mmol/l

**Combined
standard
uncertainty (%)**

Section 1 –External to manufacturer, credentialing of the Certified Reference Material

SRM917b

NIST certification of SRM917b (purity)

0.1%

Primary calibrator

SRM917b – weighed amount

Weighing procedure

0.87%

Secondary calibrator

Human Patient Specimens, e.g. Blood, Serum, Urine, CSF

Higher Order Reference Procedure – e.g. Isotope Dilution - Mass Spectrometry or Procedure of Similar Trueness and Precision

Section 2 –Internal to manufacturer, value assignment

Manufacturer's working calibrator

Manufacturer's Master Calibrator, Master Lot of Product Calibrator

Reference Procedure traceable to higher order reference procedure - e.g. Hexokinase/glucose-6-phosphate Dehydrogenase Procedure

1.21%

Product Calibrator

New Lot Commercial Product Calibrator

Procedure applying same chemistry and equipment as routine procedure, but more precisely controlled conditions and more replicates to reduce uncertainty


1.49%

Section 3 –External to manufacturer, End user's results are Traceable to Certified Reference Material and the Reference System

Commercially available system including product reagent and calibrator lots

Routine Sample – Human Patient Specimens, e.g. Blood, Serum, Urine or CSF

RESULT
Glucose in mmol/l



Proficiency Testing Scheme Results for Glucose in Serum

- Results agree with traceable reference value and acceptance criteria


GL4/06

glucose

Hexokinase Reaktion

Split 2

DGKL



| | | |
|-----------------------|-----------|-------------|
| selected participants | | 29 |
| target value | 175 | 77 |
| limits | 147 - 203 | 60,8 - 93,2 |
| mean value | 175,438 | 75,955 |
| standard deviation | 6,536 | 5,647 |