

# Standardization and Harmonization of Cardiac Troponin I Assays:

*Where have we been? Where are we now? Where are we going?*

---

Robert H. Christenson, Ph.D., DABCC, FACB  
Professor of Pathology  
Professor of Medical and Research Technology  
Director, Core Laboratory  
University of Maryland School of Medicine

# Cardiac Troponin is the Cornerstone

---

## NACB Clinical Guidelines for ACS *2007 Clin Chem and Circulation*

### Class I

- **Cardiac troponin is the preferred marker for the diagnosis of MI.** CK-MB by mass assay is an acceptable alternative when cardiac troponin is not available (Level of Evidence: A).
- In the **presence of a clinical history suggestive of ACS**, the following are considered indicative of myocardial necrosis consistent with MI (Level of Evidence: C):
  - Maximal concentration of **cardiac troponin exceeding the 99th percentile** of values (with optimal precision defined by total CV $\leq$ 10%) **for a reference control group on at least one occasion during the first 24 hours after the clinical event** (Observation of a rise and fall in values is useful in discriminating the timing of injury).

# Class 1A Recommendation

*Circulation 2002;106:1893-1900*

---

- An early invasive strategy in patients with UA/NSTEMI without serious comorbidity and who have any of the following high-risk indicators:
  - Recurrent angina/ischemia at rest or with low level activities despite intensive anti-ischemic therapy
  - *Elevated cTnI or cTnT Elevations in cTnI (0.1 ng/mL) and cTnT (0.01 ng/mL) provides superior predictive capability.*
  - New or presumably new ST-segment depression

Where have we been?

---

# Need for standardization/harmonization?

*Clin Chem 2001;47:431-437*

---

- **10 ng/mL troponin CIT standard material, value-assigned by NIST, was measured by 13 participating cTnI assays, in duplicate:**

<b>Assay</b>	<b>Value 1</b>	<b>Value 2</b>	<b>Mean</b>
1	143.7	135.7	139.7
2	46.9	47.9	47.4
3	19.9	19.9	19.9
4	49.2	50.1	49.7
5	4.3	4.1	4.2
6	12.5	12.6	12.6
7	7.4	8.9	8.2
8	6.0	6.1	6.0
9	16.5	17.4	17.0
10	12.6	13.2	12.9
11	12.6	12.4	12.5
12	18.3	18.9	18.6
13	17.6	16.9	17.3

**Mean: 28.1**  
**SD: 36.4**  
**CV: 130%**

- **Within assay reproducibility good, but measurement of 10 ng/mL material yielded results that were >30-fold different!**

# AACC cTnI Standardization Committee

---

- Desired endpoint is reducing inter-method variability
- Possible strategies
  - Define common reference material
  - Establish consensus values for cTnI in “real” patient specimens
  - Combination of SRM and patient specimens to standardize and harmonize cTnI assays

# Strategies

---

- Phase 1
  - Obtain wide variety of candidate Standard Reference Materials (cSRM)
  - Evaluate and narrow field of candidate SRMs
- Phase 2
  - Full characterization of remaining candidate SRMs
  - Evaluate commutability and investigate harmonization
  - Select final SRM
- Phase 3
  - Obtain and archive large amount of SRM
  - Fully characterize SRM
  - Validate performance of SRM
- Phase 4
  - Develop protocol for cTnI standardization
  - Assign values to serum pools traceable to SRM 2921
  - Develop reference immunoassay for troponin I

# Phase 1

---

- Obtain wide variety of candidate Reference Materials (RM)
- Evaluate and narrow field of candidate RMs



# Candidate Reference Materials (cRMs)

---

Evaluation of 10 candidate SRMs: Human or recombinant material CIT complex, CI complex or free cTnI

Designation	cTnI form	Original matrix	Final matrix	Provider
A	CI	Liquid	Original	Spectral Diagnostics, Inc., Toronto, Canada
B	CTI	Liquid	Original	Spectral Diagnostics, Inc.
C	CI	Liquid	Original	University of Miami, Miami, FL
D	CI	Liquid then lyophilized	Reconstituted	Spectral Diagnostics, Inc.
E	CTI	Liquid then lyophilized	Reconstituted	Spectral Diagnostics, Inc.
F	Free I	Lyophilized	Reconstituted	HyTest LTD, Turku, Finland
G	CTI	Lyophilized	Reconstituted	HyTest LTD
H	CTI	Liquid	Original	University of Miami
I	CI	Lyophilized	Reconstituted	University of Miami
J	CTI	Lyophilized	Reconstituted	University of Miami

# *Round Robin 1*

## cTnI Measurement Methods

---

–Evaluated 13 troponin I assays from US, Japan, Germany, the UK, and France

### **cTnI measurement system**

Access, 1st Generation

Access, 2nd Generation

ACS:180

AIA

AxSYM

Centaur

Dimension RxL

Immuno 1

Liaison

Opus, 2nd Generation

Stratus II

Triage

Vitros ECI

### **Manufacturer**

Beckman Instruments, Inc.

Beckman Instruments, Inc.

Bayer Corporation

TOSOH

Abbott Diagnostics, Inc.

Bayer Corporation

Dade Behring

Bayer Corporation

Byk-Sangtec Diagnostica

Dade Behring

Dade Behring

Biosite Diagnostics, Inc.

Ortho Clinical Diagnostics

# Conclusion: Round Robin 1

*Clin Chem 2001;47:431-437.*

---

**The outcome of this study was that to identify two cRMs that demonstrated the best performance.**

# Phase 2:

Results of AACC cTnI Standardization Committee

---

- Evaluate two candidate RMs from Phase 1
  - TIC complex (Human)
  - IC complex (Recombinant)
- Full characterization of candidate SRMs
- 15 different cTnI assays
  - Evaluate commutability
  - Investigate harmonization strategies
- Select final Reference Material

# Materials for Phase 2 Study

---


## Candidate Standard Reference Materials (cSRMs)

- TIC complex,  $\text{cSRM}_{\text{TIC}}$
- IC complex,  $\text{cSRM}_{\text{IC}}$

Characterized at NIST for purity and concentration.

cTnI-Negative Serum Pool

## LC/MS of Cardiac Troponin Complex used in SRM 2921



# Six Human Serum Pools

---

- Pool 1, ~0.4 ng/mL
- Pool 2, ~1.5 ng/mL
- Pool 3, ~15.0 ng/mL
- Pool 4, ~0.4 ng/mL
- Pool 5, ~1.5 ng/mL
- Pool 6, ~15.0 ng/mL

# cTnI 2nd Round Robin Participants and Measurement Systems

---

<u>cTnI Measurement System (15)</u>	<u>Manufacturer (10)</u>
Access	Beckman Coulter Instruments
ACS 180	Bayer Corporation
AIA	TOSOH
Alpha Dx	First Medical Inc
AxSYM	Abbott Diagnostics
Centaur	Bayer Corporation
Dimension RxL	Dade Behring
Immulite 1000	Diagnostic Products Corporation
Immulite 2000	Diagnostic Products Corporation
Immuno 1	Bayer Corporation
Liaison	Byk-Sangtec Diagnostica
Opus, 2 <sup>nd</sup> Generation	Dade Behring
Stratus CS	Dade Behring
Vidas	BioMerieux
Vitros ECI	Ortho Clinical Diagnostics



# Commutability Definition:

---

*“degree to which a material yields the same numerical relationships between results of measurements by a given set of measurement procedures applied to those types of material for which the procedures are intended.”*

European Committee for Standardization (CEN). Draft International Standard ISO/DIS 17511. ISO Central Secretariat, Brussels, Belgium, 2000.

# Commutability Assessment

---

- Calibrate systems with field cTnI methods
- Measure
  - Pools 1-6
  - 1, 5 and 10 ng/mL cSRM samples prepared in negative serum.
- Compare the pool values for each individual cTnI method to each of the other methods
- Calculate regression parameters for each method comparison
  - Method difference for pools (natural materials) determined
- Determine the method difference for the cSRM concentrations
- Compare data from serum pools (natural material) to that of the SRMS with +3/-3 SD criteria
- **Commutability graded as**
  - **YES (1)**
  - **NO (0)**

# Degree of Commutability for SRM

A.

		cRM <sub>CR</sub> in serum															Number of Comparisons Commutable
		A	B	C	D	E	F	G	H	I	J	K	L	M	N	O	
c T N I M E T H O D S	A	-	1	0	0	0	1	1	0	0	0	0	1	1	0	0	5
	B		-	1	0	1	1	1	1	1	1	1	0	1	1	1	11
	C			-	0	0	0	0	0	0	0	0	0	0	0	0	0
	D				-	1	1	0	1	1	1	1	0	0	1	1	8
	E					-	1	0	1	0	0	0	0	0	0	0	2
	F						-	1	1	1	1	0	0	1	1	0	6
	G							-	1	1	1	1	0	0	0	1	5
	H								-	1	0	0	0	1	0	0	2
	I									-	0	0	0	1	0	0	1
	J										-	1	0	1	0	1	3
	K											-	1	1	1	1	4
	L												-	0	0	0	0
	M													-	0	0	0
	N														-	0	0
	O															-	0

**Degree of Commutability is 47 of 105 (45%)**

B.

		cRM <sub>CI</sub> in serum															Number of Comparisons Commutable
		A	B	C	D	E	F	G	H	I	J	K	L	M	N	O	
c T N I M E T H O D S	A	-	0	0	0	1	0	1	0	0	1	1	0	0	0	1	5
	B		-	0	0	0	0	0	0	0	0	0	0	0	0	0	0
	C			-	0	0	0	1	0	0	1	0	1	1	0	0	4
	D				-	1	0	0	1	1	1	1	0	0	1	1	7
	E					-	0	1	1	0	1	1	0	0	0	1	5
	F						-	0	0	0	0	0	0	0	0	0	0
	G							-	1	0	1	0	1	0	0	0	3
	H								-	1	1	1	0	1	0	1	5
	I									-	1	1	0	0	0	1	3
	J										-	1	0	1	0	0	2
	K											-	1	1	1	1	4
	L												-	1	1	0	2
	M													-	1	0	1
	N														-	0	0
	O															-	0

**Degree of Commutability is 41 of 105 (39%)**

# Harmonization

---

- Agreement between cTnI measurements performed by different methods in the biological matrix (patient serum).


# Harmonization: Present State

---

- Systems calibrated with manufacturers' calibrators and instructions
- Systems calibrated with cSRM
- 6 serum pools measured with each system
- Determine inter-method variability

# AACC cTnI Standardization Phase 2

Use of SRM 2921 as the common calibrator did not improve cTnI standardization




# Harmonization by Alignment

---

- Calibrate methods (field and cRMs)
- Measure 6 serum pools with each system
- Designate “common comparison” system
- Regression analysis to yield parameters
- Align each serum pool with regression parameters
- Determine inter-method variability of aligned results for each pool

# Harminization of cTnI Measurements is Possible

---





## **Certification of SRM 2921**

- Quantification of troponin I in the troponin TIC complex performed using two methods, LC coupled to amino acid analysis (calibrated using SRM 2389 for SI-traceability) and LC/UV using a purified troponin I primary standard.
- Extensive structural characterization performed using LC/MS for molecular mass and structural heterogeneity evaluation of the intact protein molecules and MALDI/MS on peptide digests of the troponin T, troponin I, and troponin C subunits



Where are we now?

---

# Phase 3

---

- ***Standard Reference Material 2921***
  - Vials were obtained from NIST, Gaithersburg, MD.
- ***Patient Pools***
  - Eight (8) patient pools were prepared at various cTnI concentrations in accordance with CLIS guidances.
- ***Recovery: Standard Addition Experiment***
  - cTnI positive serum pools were spiked with RM 2921 to gravimetrically produce SRM 2921 target values of 0.5, 1, 2, and 4 ng/mL of cTnI.
- ***Strategy for Stabilizing cTnI Assays with SRM 2921***
  - Characteristic Factor Determined for each cTnI method
  - Validation of RM 2921 Characteristic Factors

# Methods Included in Phase 3

---

## Measurement System

Access 2

ACS 180

ADVIA Centaur

ARCHITECH System 135

AxSYM cTnI ADV

Dimension RxL

IMMULITE

IMMULITE Turbo

Stratus CS

Vidas

Vitros ECi / ECiQ

## Manufacturer

Beckman Instruments, Inc

Bayer Corporation

Bayer Corporation

Abbott Diagnostics, Inc.

Abbott Diagnostics, Inc.

Dade Behring, Inc.

Diagnostic Products Corp

Diagnostic Products Corp


Dade Behring, Inc.

BioMerieux

Ortho Clinical Diagnostics

# Scatterplot of the 8 serum pools assayed by field methods

---




# After Harmonization

---

<b>cTnI</b>	<b>Pool 1</b>	<b>Pool 2</b>	<b>Pool 3</b>	<b>Pool 4</b>	<b>Pool 5</b>	<b>Pool 6</b>	<b>Pool 7</b>	<b>Pool 8</b>
<b>Assay</b>	<b>ng/mL</b>	<b>ng/mL</b>	<b>ng/mL</b>	<b>ng/mL</b>	<b>ng/mL</b>	<b>ng/mL</b>	<b>ng/mL</b>	<b>ng/mL</b>
1	0.08	0.18	0.24	0.42	0.87	2.48	2.48	7.38
2	0.09	0.15	0.25	0.42	0.93	2.52	3.53	7.44
3	0.11	0.17	0.26	0.45	0.95	2.43	3.37	7.53
4	0.04	0.15	0.32	0.47	0.86	2.76	3.33	7.44
5	0.07	0.16	0.25	0.44	0.92	2.59	3.47	7.44
6	0.11	0.20	0.26	0.41	0.78	2.40	3.68	7.42
7	0.00	0.09	0.23	0.45	1.22	2.70	3.19	7.45
8	0.05	0.12	0.23	0.43	1.03	2.77	3.21	7.46
9	0.08	0.18	0.27	0.45	0.84	2.45	3.59	7.44
10	NA	NA	0.47	0.59	0.75	1.97	3.90	7.38
11	0.12	0.20	0.27	0.42	0.92	2.59	3.16	7.56
<b>Mean</b>	<b>0.07</b>	<b>0.16</b>	<b>0.28</b>	<b>0.45</b>	<b>0.91</b>	<b>2.52</b>	<b>3.47</b>	<b>7.45</b>
<b>CV</b>	<b>69%</b>	<b>21%</b>	<b>24%</b>	<b>11%</b>	<b>14%</b>	<b>9%</b>	<b>7%</b>	<b>1%</b>

# Variability Across cTnI Methods

---





# Assure Stability of cTnI Methods


---

- Characteristic Factor
  - Analytical response of methods to SRM 2921 concentrations
- Calculation:
  - SRM 2921 theoretical value/ measured value
- Characteristic Factor should be a constant

# Characteristic Factor (CF) for SRM 2921

## cTnI Methods

	1	2	3	4	5	6	7	8	9	10	11
<i>CF</i>	<b>1.67</b>	<b>5.99</b>	<b>6.11</b>	<b>2.96</b>	<b>1.64</b>	<b>2.69</b>	<b>4.52</b>	<b>3.61</b>	<b>3.42</b>	<b>0.57</b>	<b>1.70</b>
<i>Validation</i>	<b>1.65</b>	<b>6.07</b>	<b>6.09</b>	<b>3.09</b>	<b>1.61</b>	<b>2.75</b>	<b>3.79</b>	<b>3.33</b>	<b>3.51</b>	<b>0.61</b>	<b>1.90</b>




Where are we going?

---

# Metabolism: cTnI is Heterogeneous Analyte

*Circulation 2000;102;1221-1226*

Western Blot: Reactivity with cTnI Antibody



# Phase 4

---

**Goal:** To improve standardization of clinical cTnI measurement through the development of a serum-based reference material

## **Tasks:**

- Development of reference measurement procedure for serum cTnI
- Produce pooled serum reference materials (multi-level)
- Evaluate commutability of reference materials and effectiveness to improve inter-assay comparability
- JCTLM submission and approval

# Development of a Reference Measurement Procedure

---

**Objective:** investigate the use of a critically-evaluated immunoassay as a higher-order measurement procedure

- Immunoassay development a collaboration between NIST(USA) and NPL (UK) with guidance on cTnI antibodies by HyTest Ltd.
- Immunoassay calibrated using NIST SRM 2921
- Clinical expertise and assay evaluation samples provided by the IFCC's Committee for the Standardization of Markers of Cardiac Damage (C-SMCD)
- Interlaboratory evaluation of the cTnI immunoassay to be undertaken as a pilot project of the Bioanalysis Working Group of the CCQM

# *Summary*

*Where have we been? Where are we now? Where are we going?*

---

- cTnI showed ~30-fold difference in methods
- Candidate Reference Materials examined
- SRM 2921 characterized
- Commutability issues: ~50% of methods
- Harmonization possible using serum pools
- cTnI is a heterogeneous analyte
- International effort for standardization using serum pools traceable to SRM 2921

# Thank you!

rchristenson@umm.edu

---

