

February 23, 2024

# Technical Protocol for Regional Key Comparison on LS2P Microphone Calibration in Pressure Field - SIM.AUV.A-K6

## Amendment 2

Page 5, Section “**Timetable**”

Replace by

### **Timetable**

The timetable of the comparison is presented in Table 2. It is essential that the microphones be transported to the next participant on time. Any change of the timetable must be justified and agreed.

Table 2 – Timetable for the SIM.AUV.A-K6 comparison.

1) Check of microphones stability	
National Metrology Institute	Period
INMETRO	May 29 to September 08, 2023
2) Circulation of the microphones	
National Metrology Institutes	Period
INMETRO	September 11 to September 29, 2023
Transportation of the microphones by previous participant to the next one	
NRC	February 12 to March 1, 2024
Transportation of the microphones by previous participant to the next one	
NIST	March 25 to April 12, 2024
Transportation of the microphones by previous participant to the next one	
CENAM	May 6 to May 24, 2024
Transportation of the microphones by previous participant to the next one	
INMETRO	June 17 to July 5, 2024
Transportation of the microphones by previous participant to the next one	
LACOMET	July 29 to August 16, 2024
Transportation of the microphones by previous participant to the next one	
INTI	September 9 to September 27, 2024
Transportation of the microphones by previous participant to the next one	
INMETRO	October 21 to November 8, 2024

(continue)

3) Analysis of the reported results	
National Metrology Institute	Period
INMETRO	November 11, 2024 to January 24, 2025

Notes:

1. A period of three weeks is allocated for the microphones calibration by each participant.
2. A period of three weeks is also allocated for the transportation of the microphones between participants. Because of the difference between customs formalities of each country, it is difficult to know how long it actually takes to clear the goods. Therefore, each participant shall consider the transportation as early as possible within the period in order to avoid any delays due to customs. It is highly recommended that each participant get in contact with the next one in advance in order to check all specific customs requirements/procedures to be fulfilled, assure a expedite export/import process and avoid unnecessary delays with the transference of the microphones.
3. INMETRO will carry out measurements in **three periods** during the circulation of the microphones to monitor their stability, but only the results obtained during the first period will be used for comparison with the reference values.

The period from **December 23, 2024 to January 3, 2025** is assumed a recess period due to the holidays of Christmas and New.

---